

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526
Name		Joseph V. SHOEMAKER	ID# 4188 Date 3-8-12
A	Agency	SAND POINT PD - SPD 1	907-383-3700
Instrument Location		SAND POINT PD, SAND POINT, AK 99661	
Alco S/N		4 W	Target Value .080 High Pressure 700
Alco Test Value Average		.076	.077
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		Joseph V. Shoemaker	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012.

Nikki Roth  
Nikki Roth, Notary Public  
My Commission Expires With Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130526

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

MARCH 07, 2012

OPERATOR'S NAME:

SHOENAKER

OPERATOR'S NUMBER: 4188

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: SDPI

CASE/REPORT: K

TEST TYPE: V

ALCO TARGET VALUE: .080

ALCO S/N: 4W

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.22 in

ALCO TARGET .080 12:59

BLANK TEST .080 13:00

INTERNAL STANDARD VERIFIED 13:00

ALCO TV 30.22 in .076 13:02

BLANK TEST .080 13:03

SUBJECT SAMPLE .080 13:04

BLANK TEST .080 13:05

ALCO TV 30.22 in .077 13:05

BLANK TEST .080 13:06

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

MARCH 07, 2012

TIME 13:07

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 42c

BAROMETER: 30.24 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@ABCDEF6

HIJKLMNPOQRSTUVWXYZ[\]^\_`abcde fghijklmno

pqrstuvwxyz{ }+~



BT9 03/00